

Resultados de Biometria Hemática Ciclo 04-2012 Programa EEXCEL

MEDIA	13.31	38.46	7,829.5	4,600,909	29.41	33.80	85.12	274,974	
DS	3.31	11.71	2,912.5	1,158,820	4.49	8.38	20.82	97,561	
CV	24.89	30.44	37.20	25.19	15.26	24.79	24.46	35.48	
Criterio	B±7%	B±6%	B±15%	B±6%	B±6%	B±6%	B±6%	B±25%	
Valor ±	0.93	2.31	1174.4	276055	1.76	2.03	5.11	68744	
LABORAT	HB	HTO	LUECOS	ERIS	HGM	CMHG	VGM	PLAQ	PROCESÓ
E-007	14.1	41.0	7,520	5,010,000	28.1	34.4	81.8	260,000	Automatizado
E-010	13.4	39.3	6,800	4,800,000	28.0	34.0	82.0	273,000	Automatizado
E-011	13.8	42.1	8,400	4,680,000	90.0	29.5	32.8	317,000	Automatizado
E-013	12.9	33.3	7,700	4,300,000	29.9	38.7	77.1	287,000	Automatizado
E-014	12.4	38.0	8,000	4,180,000	29.6	32.6	90.9	256,000	Automatizado
E-017	13.6	38.7	7,800	4,600,000	29.4	35.1	83.5	290,000	Automatizado
E-018	13.7	41.4	7,500	4,900,000	27.9	33.0	84.6	222,000	Automatizado
E-027	15.1	42.0	7,900	4,590,000	32.8	35.9	91.6	263,000	Automatizado
E-028	13.4	39.3	8,200	4,620,000	29.0	34.1	85.1	264,000	Automatizado
E-033	14.7	35.7	7,800	4,330,000	33.9	41.2	82.3	260,000	Automatizado
E-035	13.0	38.9	8,000	5,100,000	28.6	33.5	85.2	292,000	Automatizado
E-038	12.4	34.1	8,100	4,300,000	29.1	36.3	80.2	200,000	Automatizado
E-039	13.3	40.0	8,100	4,570,000	29.1	33.2	87.7	338,000	Automatizado
E-040	13.3	37.0	7,600	4,500,000	28.8	35.6	81.0	260,000	Automatizado
E-043	13.5	38.3	7,900	4,320,000	31.2	35.2	88.7	247,000	Automatizado
E-044	14.6	42.9	8,100	4,710,000	30.9	34.0	91.2	295,000	Automatizado
E-045	13.6	38.8	7,900	4,520,000	30.0	35.0	85.9	298,000	Automatizado
E-046	14.2	41.2	7,500	4,600,000	30.8	34.4	89.6	245,000	Automatizado
E-048	14.2	38.6	8,000	4,910,000	28.9	36.8	79.0	192,000	Automatizado
E-049	14.2	39.4	7,900	5,010,000	28.3	35.9	78.7	361,000	Automatizado
E-050	12.9	38.0	7,600	4,700,000	27.5	34	81.0	289,000	Automatizado
E-051	13.7	38.8	7,900	4,500,000	30.5	35.4	86.1	277,000	Automatizado
E-059	13.6	39.0	7,600	4,600,000	29.3	34.9	84.1	301,000	Automatizado
E-063	13.2	38.4	7,100	4,500,000	29.3	34.4	85.1	309,000	Automatizado
E-064	12.9	35.9	7,500	4,600,000	28.4	36.6	74.0	299,000	Automatizado
E-065	16.0	42.0	8,600	4,900,000	32.7	38.1	85.8	275,000	Automatizado
E-070	13.8	41.9	8,000	4,800,000	29.3	33.6	88.2	272,000	Automatizado
E-073	13.8	40.6	3,400	4,700,000	29.6	33.9	87.4	271,000	Automatizado
E-077	16.2	47.2	10,300	5,480,000	29.5	34.3	86.1	259,000	Automatizado
E-079	10.6	35.0	8,400	3,800,000	27.8	30.5	89.0	330,000	Automatizado
E-080	13.8	38.3	8,210	4,420,000	31.0	35.0	87.0	265,000	Automatizado
E-081	13.4	42.0	8,260	4,580,000	29.7	31.9	93.3	291,000	Automatizado
E-083	13.9	37.4	7,100	4,700,000	29.8	37.2	80.3	225,000	Automatizado
E-087	14.5	37.2	7,300	4,700,000	31.0	33.5	80.0	488,000	Automatizado
E-097	13.8	35.7	7,900	4,500,000	30.9	38.5	80.1	226,000	Automatizado
E-108	13.3	34.3	7,500	4,470,000	29.7	38.6	77.0	259,000	Automatizado
E-111	13.5	34.6	7,700	4,600,000	29.2	39.1	75.0	300,000	Automatizado
E-116	13.3	39.9	7,900	4,700,000	28.1	33.3	84.3	319,000	Automatizado
E-117	14.2	44.2	9,400	5,000,000	28.3	32.1	88.4	281,000	Automatizado
E-118	14.1	40.7	8,000	4,700,000	29.7	34.5	85.9	310,000	Automatizado
E-119	15.4	44.2	9,700	5,000,000	30.3	34.8	87.1	271,000	Automatizado
E-130	14.2	41.7	8,300	4,730,000	29.9	34.0	81.1	257,000	Automatizado
E-132	13.3	34.3	7,500	4,470,000	29.7	38.6	77.0	259,000	Automatizado

Resultados de Biometria Hemática Ciclo 04-2012 Programa EEXCEL

MEDIA	13.31	38.46	7,829.5	4,600,909	29.41	33.80	85.12	274,974	
DS	3.31	11.71	2,912.5	1,158,820	4.49	8.38	20.82	97,561	
CV	24.89	30.44	37.20	25.19	15.26	24.79	24.46	35.48	
Criterio	B±7%	B±6%	B±15%	B±6%	B±6%	B±6%	B±6%	B±25%	
Valor ±	0.93	2.31	1174.4	276055	1.76	2.03	5.11	68744	
LABORAT	HB	HTO	LUECOS	ERIS	HGM	CMHG	VGM	PLAQ	PROCESÓ
E-003	13.7	42.0		4,620,000	29.5	32.5	91.0		Manual
E-009	13.0	36.0		3,800,000	34.2	36.1	97.7	250,000	Manual
E-020	11.5	36.0		4,100,000		31.9	87.2		Manual
E-021	11.5	36.0		4,100,000	28.0	31.9	87.8	341,666	Manual
E-029	13.3	40.0		4,300,000	33.2	30.0	93.0		Manual
E-058	11.0	36.0		3,800,000		30.0	94.0		manual
E-060	11.5	35.0		3,760,000	30.0	33.0	93.0		Manual
E-066	11.2	33.6		3,700,000	30.3	33.3	90.8		Manual
E-072	11.4	34.4		3,800,000	30.0	33.1	90.5		Manual
E-090	10.3	32.0				30.0	88.0		Manual
E-092	10.2	32.0	3,620			30.0	88.0		Manual

Excelentes	0 a 0.5
Buenos	>0.5 a 1.0
Satisfactorio	>1.0 a 2.0
No aceptables	> 2.0