

Resultados de Biometria Hemática Ciclo02-2012 Programa EEXCEL

MEDIA	5.90	17.41	1,966.7	2,338,293	25.44	33.31	75.72	82,816	
DS	1.21	4.76	671.0	745,340	8.67	9.12	12.09	26,851	
CV	20.57	27.33	34.12	31.88	34.07	27.39	15.96	32.42	
Criterio	B±7%	B±6%	B±15%	B±6%	B±6%	B±6%	B±6%	B±25%	
Valor ±	0.41	1.04	295.0	140298	1.53	2.00	4.54	20704	
LABORAT	HB	HTO	LUECOS	ERIS	HGM	CMHG	VGM	PLAQ	PROCESÓ
E-007	6.04	17.98	1810.00	2,530,000	23.84	33.56	71.04	85000.00	Automatizado
E-010	5.00	16.80	1600.00	2,320,000	24.10	33.30	73.00	89000.00	Automatizado
E-011	6.30	19.80	2400.00	2,420,000	81.80	26.00	31.80	135000.00	Automatizado
E-013	6.00	16.70	2100.00	2,300,000	26.10	35.90	72.60	116000.00	Automatizado
E-017	6.10	18.10	2100.00	2,490,000	24.50	33.70	72.70	86000.00	Automatizado
E-018	5.80	18.20	1600.00	2,430,000	23.80	31.80	75.10	77000.00	Automatizado
E-027	6.40	18.90	2000.00	2,320,000	27.50	33.80	81.70	79000.00	Automatizado
E-028	5.70	17.40	2100.00	2,300,000	24.70	32.50	75.80	64000.00	Automatizado
E-033	6.20	16.20	2200.00	2,220,000	28.10	38.60	72.80	74000.00	Automatizado
E-035	5.50	17.90	2000.00	2,330,000	23.80	31.00	76.60	78000.00	Automatizado
E-038	5.30	15.10	1900.00	2,150,000	24.60	35.00	70.60	63000.00	Automatizado
E-039	5.50	18.40	1900.00	2,320,000	23.70	29.80	79.50	188000.00	Automatizado
E-043	5.70	16.50	1800.00	2,130,000	26.70	34.50	77.50	94000.00	Automatizado
E-044	6.20	18.80	1900.00	2,420,000	25.60	32.90	77.90	95000.00	Automatizado
E-045	6.20	18.20	2100.00	2,360,000	26.20	34.00	77.40	113000.00	Automatizado
E-046	6.00	19.10	1800.00	3,370,000	25.30	31.40	80.80	82000.00	Automatizado
E-048	6.30	17.30	2000.00	2,520,000	25.10	36.10	69.00	58000.00	Automatizado
E-049	6.00	17.20	1900.00	2,440,000	24.60	34.90	70.70	91000.00	Automatizado
E-050	5.50	16.70	2000.00	2,330,000	23.50	32.90	71.00	88000.00	Automatizado
E-051	5.90	18.40	2000.00	2,340,000	25.30	32.20	78.50	89000.00	Automatizado
E-059	5.80	17.60	2900.00	2,300,000	24.70	32.90	75.20	81000.00	Automatizado
E-063	5.60	17.20	1800.00	2,270,000	24.80	32.80	75.70	87000.00	Automatizado
E-065	6.50	18.60	2300.00	2,420,000	26.80	34.80	77.00	75000.00	Automatizado
E-070	6.10	20.23	2120.00	2,500,000	24.40	30.15	79.68	98000.00	Automatizado
E-073	6.01	19.10	1010.00	2,430,000	24.70	31.40	78.60	86500.00	automatizado
E-077	6.80	21.20	4900.00	2,760,000	24.60	32.10	76.60	57000.00	Automatizado
E-079	4.30	16.00	3500.00	2,100,000	20.00	27.00	76.00	135000.00	automatizado
E-080	6.10	18.00	2120.00	2,300,000	26.10	33.60	76.00	91000.00	Automatizado
E-081	6.00	19.00	1950.00	2,280,000	27.20	31.80	86.30	74000.00	Automatizado
E-083	6.70	17.60	1700.00	2,470,000	27.10	38.00	71.00	81000.00	Automatizado
E-087	6.00	16.20	1800.00	2,300,000	26.30	30.00	70.00	81000.00	Automatizado
E-097	4.70	16.60	1900.00	2,360,000	20.00	28.40	70.30	70000.00	Automatizado
E-111	6.10	16.50	1900.00	2,400,000	25.20	37.10	68.00	97000.00	Automatizado
E-116	5.70	18.30	2000.00	2,420,000	23.50	31.10	75.50	87000.00	Automatizado
E-117	5.90	18.90	2100.00	2,480,000	23.80	31.20	76.20	78000.00	Automatizado
E-118	6.00	17.30	1900.00	2,330,000	25.70	34.50	74.40	85000.00	Automatizado
E-119	6.50	19.40	3800.00	2,500,000	26.00	33.70	77.10	70000.00	Automatizado
E-130	6.20	18.20	2000.00	2,400,000	25.80	33.80	76.20	61000.00	Automatizado
E-132	5.90	15.60	1900.00	2,300,000	25.70	38.00	68.00	103000.00	Automatizado
E-014	6.10	17.00	3400.00	1,870,000	32.60	35.80	90.90	112000.00	Automatizado
E-040	6.00	17.00	2000.00	2,410,000	25.00	34.50	72.00	77000.00	Automatizado
E-064	6.40	20.00	2100.00	3,050,000	20.80	31.70	66.00	91000.00	Automatizado

Resultados de Biometría Hemática Ciclo02-2012 Programa EEXCEL

MEDIA	5.90	17.41	1,966.7	2,338,293	25.44	33.31	75.72	82,816	
DS	1.21	4.76	671.0	745,340	8.67	9.12	12.09	26,851	
CV	20.57	27.33	34.12	31.88	34.07	27.39	15.96	32.42	
Criterio	B±7%	B±6%	B±15%	B±6%	B±6%	B±6%	B±6%	B±25%	
Valor ±	0.41	1.04	295.0	140298	1.53	2.00	4.54	20704	
LABORAT	HB	HTO	LUECOS	ERIS	HGM	CMHG	VGM	PLAQ	PROCESÓ
E-003	4.96	16.00		1,760,000	28.00	31.00	90.00		Manual
E-009	6.70	18.00		1,900,000	35.30	37.20	94.70	87000.00	Manual
E-020	5.40	17.00		1,900,000		31.70	89.40		Manual
E-021	5.10	16.00		1,800,000	28.30	31.80	88.88	86893.00	Manual
E-029	6.00	18.00		2,100,000	33.30	30.00	85.70		Manual
E-058	5.90	18.00		1,900,000		33.00	94.00		Manual
E-060	5.28	16.00		2,000,000	26.40	33.00	80.00		Manual
E-066	5.00	15.10		1,600,000	31.30	33.10	94.40		Manual
E-072	5.60	17.00		1,800,000	31.10	32.90	94.40		Manual
E-090	5.60	17.00				18.00	80.00		Manual
E-092	5.40	17.00		2,130,000		18.00	80.00		Manual

Excelentes	0 a 0.5
Buenos	>0.5 a 1.0
Satisfactorio	>1.0 a 2.0
No aceptables	> 2.0