

Resultados de Química Clínica Ciclo 02-04 de 2012, Programa EEXCEL

MEDIA	8.92	131.16	2.78	136.57	2.06	4.81	11.78	117.48	266.58	4.11	338.30	6.51	260.82	6.02	4.70	159.71	245.67	110.72
DS	3.45	48.69	0.95	51.58	0.92	1.79	5.19	38.12	78.69	1.25	177.12	1.67	96.02	0.32	1.68	37.75	93.14	45.40
CV	38.64	37.12	34.00	37.77	44.82	37.16	44.09	32.45	29.52	30.48	52.36	25.69	36.81	5.31	35.71	23.64	37.91	41.00
Criterio	B ±17%	B±20%	B±20%	B±20%	B±20%	B± 20 %	±1mg/L	B±5%	B±10%	B± 15 %	B±30%	B±30%	B±10%	±0.5mmol	B±10%	±4mmol/	B±25%	B±9%
Valor ±	1.5	26.2	0.6	27.3	0.4	1.0	1.0	5.9	26.7	0.6	101.5	2.0	26.1	0.5	0.5	4.0	61.4	10.0
LAB	A.URICO	ALAT	ALB	ASAT	BIL. D	BIL. T.	CALCIO	CLOR	COLEST	CREAT	FOS.ALC	FOSFATO	GLUC	POTASIO	ROT.TOT/	SODIO	TRIG	UREA
E-003	9.80	136.00		137.00	2.04	4.81			240.00				209.00				249.00	101.00
E-007	9.50	140.00	3.20	145.00	2.09	5.95	19.00	121.00	243.00	4.10	275.00	6.20	277.00	6.10	5.00	165.00	232.00	110.00
E-009	9.00								295.00	3.70			230.00				240.00	84.80
E-010	14.20	125.00	2.30	128.00	2.09	5.49	8.50		283.40	3.48	468.00		172.60		5.70		186.00	103.00
E-011	8.20								257.00	3.40			380.00				205.00	120.00
E-013	8.70	154.00	2.50	177.00		3.30	11.70	115.00	251.00	4.60	302.00	6.60	269.00	5.90	4.50	159.00	299.00	
E-014	7.08	143.40	4.29	159.10	2.25	6.39			229.05	3.70	569.90		240.57		3.54		159.13	185.40
E-017	8.10	139.00	2.90	145.00	2.10	5.65	10.56	112.00	243.00	4.00	550.00	6.00	275.00	6.30	4.50	161.00	215.00	120.00
E-018	7.50	135.20	2.80	115.60	2.40	5.20	12.40	119.00	278.80	3.30	453.00		279.00	5.90	4.70	156.00	235.00	118.80
E-020	9.50	124.00	2.60	127.00	2.71	5.88			252.00	5.50	454.00		263.00		4.90		238.00	
E-021	9.50	120.00	2.70	112.00	2.40	4.30			334.50	2.80	122.40		235.50		5.70		341.30	78.50
E-027	9.00	139.00	2.70	143.00	1.13	4.66		116.00	273.00	3.91	252.00	6.40	254.60	5.98	4.40	158.90	214.90	121.60
E-028	8.90	138.00	3.10	132.00	3.42	5.27			264.00	3.89	505.00	6.80	255.00		4.90		256.00	112.00
E-029	9.80								305.00	6.30			267.00				269.00	2.10
E-033	9.40	116.30	2.10	126.30	2.20	4.70			269.80	5.20	387.00		221.00		4.70		215.40	72.00
E-035	8.80		3.20				10.20	118.90	271.00	3.92		7.20	292.00	6.27	5.10	158.30	231.00	96.00
E-038	8.50	123.00	2.40	168.00		3.20	12.10	119.00	262.00	4.50		5.70	271.00	5.70	4.20	159.00	297.00	106.00
E-039	1.50		3.30		2.10	5.80			245.00	4.00			250.00		4.80		240.00	140.50
E-040	8.10								256.00	3.75			252.00				226.00	104.00
E-043	8.20	218.30	2.20	107.00	1.90	2.70	12.70	103.00	277.00	4.50	232.00		268.00	5.20	3.80	159.00	285.00	101.00
E-044	9.70	132.70	3.30	156.20	2.00	4.70			294.50	4.10	359.90		254.40		4.50		245.10	121.10
E-045	10.10	115.00		113.30					229.80	3.80			246.60				239.90	116.10
E-046	6.90	113.00	3.80	137.00	1.70	4.80	8.70	123.90	224.80	4.10	482.00	6.90	299.60	6.10	7.40	148.90	188.20	128.40
E-048	9.60	112.00	3.40	91.00	2.86	5.38			360.00	4.30	580.00		260.00		4.90		355.00	139.10
E-049	10.00	115.00	3.00	110.00	1.37	5.84	15.40	122.60	319.60	3.89	365.00	5.80	273.50	6.14	5.63	158.90	225.90	131.75
E-050	8.90	10.00		126.00	2.25	4.68			311.00	4.00			329.00				314.00	160.70
E-051	8.40	154.00	2.40	168.00					232.00	4.50	297.00		259.00				299.00	102.70
E-058	4.50								208.00	4.20			190.00				173.00	90.00
E-059	8.80	144.00		150.00		2.90	11.50	116.00	242.00	4.70	252.00	4.90	276.00	6.10	4.20	165.00	282.00	99.00

Resultados de Química Clínica Ciclo 02-04 de 2012, Programa EEXCEL

MEDIA	8.92	131.16	2.78	136.57	2.06	4.81	11.78	117.48	266.58	4.11	338.30	6.51	260.82	6.02	4.70	159.71	245.67	110.72
DS	3.45	48.69	0.95	51.58	0.92	1.79	5.19	38.12	78.69	1.25	177.12	1.67	96.02	0.32	1.68	37.75	93.14	45.40
CV	38.64	37.12	34.00	37.77	44.82	37.16	44.09	32.45	29.52	30.48	52.36	25.69	36.81	5.31	35.71	23.64	37.91	41.00
Criterio	B ±17%	B±20%	B±20%	B±20%	B±20%	B± 20 %	±1mg/L	B±5%	B±10%	B± 15 %	B±30%	B±30%	B±10%	±0.5mmol	B±10%	±4mmol/	B±25%	B±9%
Valor ±	1.5	26.2	0.6	27.3	0.4	1.0	1.0	5.9	26.7	0.6	101.5	2.0	26.1	0.5	0.5	4.0	61.4	10.0
LAB	A.URICO	ALAT	ALB	ASAT	BIL. D	BIL. T.	CALCIO	CLOR	COLEST	CREAT	FOS.ALC	FOSFATO	GLUC	POTASIO	ROT.TOT/	SODIO	TRIG	UREA
E-060	8.30	119.00	5.40	133.00	1.59	2.88			257.00	3.80	666.00		295.00		7.20		200.00	161.00
E-063	8.40	144.00	2.40	180.00		3.10	11.80	119.00	246.00	4.60	263.00	6.70	249.00	6.40	4.50	162.00	326.00	101.00
E-064	9.70	191.00		145.00	1.89	4.15	11.40	127.00	292.00	4.70	230.00	7.70	235.10	5.90		164.00	237.00	126.50
E-065	9.00	118.00	3.00	111.00	2.53	4.67			263.00	2.00			239.00		5.00		172.00	112.00
E-066	8.30								293.90				252.00				234.20	113.40
E-070	9.00	168.00	2.40	181.00	2.10	3.80	12.00	119.00	239.00	3.47	277.00	7.20	187.00	6.10	5.10	158.00	237.00	100.50
E-072	10.40								358.70	4.20			238.10				237.00	47.40
E-073	9.10	131.00		126.00	1.20	5.20			293.00	4.43	354.00		284.00				251.00	90.00
E-077	8.70	141.00	3.10	135.00	2.32	5.31	11.58	114.20	269.00	3.96	460.00	6.70	239.00	6.18	4.50	153.90	239.00	123.00
E-079	8.90								263.50	4.22			187.00				190.00	126.00
E-080	9.10	39.00	2.80	126.00	1.75	4.36			281.00	4.10	361.00		265.00		4.60		268.00	126.00
E-081	12.30								280.00	4.00			355.00				224.00	168.00
E-083	8.40	146.00	3.10	161.00	2.14	4.83	12.40	114.00	259.00	4.20	588.00	6.20	272.00	5.60	5.20	160.00	239.00	121.00
E-087	8.70								280.00	3.70			303.00				312.00	130.00
E-090	11.70	118.00		131.00	1.74	5.66			295.00	3.70	515.00		284.00		6.40		229.00	93.70
E-092	12.00	118.00		131.00	1.74	5.66			305.00	4.00	515.00		298.00		6.40		230.00	97.00
E-097	8.40	150.00		163.00	0.93	3.10		116.00	255.00	5.00	223.00	5.90	255.00	5.80	3.80	159.00	278.00	106.00
E-132	7.40		2.70				12.80	114.00	243.00	4.80	242.00	6.10	247.00	6.60	4.50	158.00	280.00	96.00

LAB
E-003
E-007
E-009
E-010
E-011
E-013
E-014
E-017
E-018
E-020
E-021
E-027
E-028
E-029
E-033
E-035
E-038
E-039
E-040
E-043
E-044
E-045
E-046
E-048
E-049
E-050
E-051
E-058
E-059

LAB
E-060
E-063
E-064
E-065
E-066
E-070
E-072
E-073
E-077
E-079
E-080
E-081
E-083
E-087
E-090
E-092
E-097
E-132